



## Final Inspection Report

## Order information

Customer name	EUROTEC DEVELOPPEMENT
Customer purchase order	CF00001464
Sales order number / Item	3011075333/0060
Internal order number / Item	3020537805000060

## Device information

Description	Micropilot FMR10
TAG	
Serial number	T300B901179
Order code	FMR10-1144/0
Extended order code	FMR10-AAQBMWDEWFE2+R7

## Additional information

Output type	4...20 mA / Bluetooth®
Software version	01.00.11
Ambient temperature / - humidity / - pressure	21.6 °C ±2 °C / 60 % rel. ±40 % rel. / 980 mbar ±30 mbar
Adjusted measuring range	E= 12000mm F= 11800mm
Maximum permissible error	±5 mm

## Test result

Hereby we confirm that all applicable tests according to the Quality plan (IP01017F) have been performed successfully.

Test	Procedure number	Test description
Contract review	TS00001F	As required in ISO 9001
Incoming goods inspection	TS00003F	Verification of conformance to the specified requirements
Test of printed circuit boards	TS00009F	In-circuit test and functional test
Functional test of instrument	TS00042F	Verification of instrument function
Calibration of instrument	TS00020F	Measurement, adjustment and verification of instrument accuracy
Outgoing visual check	TS00006F	Visual inspection of completeness and correctness of the instrument and the markings
Logistical verification	TS00005F	Automatic verification in the delivery phase via barcode that the instrument correspond to the order

Test point	Nominal value	2759.62 mm
	Measurement value	2760.04 mm
	Deviation	0.42 mm

Hereby we confirm that all measuring equipment used to assure the quality of the products has been calibrated and is traceable to national or international standards.

## Execution

Maulburg, 2022-03-04, P00007

This document was generated electronically and is valid without signature.